CERTIFICATE

Issued Date: Apr. 19, 2022
Report No. : 2220291R-E3012310001-E

This is to certify that the following designated product

Product	:	**Medical Premium Windows Touch Panel PC**
Trade name	:	**MACTRON GROUP**
Model Number	:	**WMP2157, WCP2157**
Company Name	:	**MACTRON GROUP CO., LTD.**

This product, which has been issued the test report listed as above in DEKRA Testing and Certification Co., Ltd. Laboratory, is based on a single evaluation of one sample and confirmed to comply with the requirements of the following EMC standard.

EN 60601-1-2: 2015
EN IEC 61000-3-2: 2019
EN 61000-3-3: 2013+A1: 2019

IEC 60601-1-2: 2014
IEC 61000-4-2 Ed. 2.0: 2008
IEC 61000-4-3 Ed. 4.0: 2020
IEC 61000-4-4 Ed. 3.0: 2012
IEC 61000-4-5 Ed. 3.1: 2014+A1: 2017
IEC 61000-4-6 Ed. 4.0: 2013
IEC 61000-4-8 Ed. 2.0: 2009
IEC 61000-4-11 Ed. 3.0: 2020

TEST LABORATORY

Vincent Lin / Director

DEKRA Testing and Certification Co., Ltd.
No. 5-22, Ruishukeng, Linkou Dist., New Taipei City 24451, Taiwan
TEL:+886-2-8601-3788 FAX:+886-2-8601-3789 Email: info.tw@dekra.com http://www.dekra.com.tw